

NIZAM'S INSTITUTE OF MEDICAL SCIENCES  
(A University Established Under State Act)  
PANJAGUTTA :: HYDERABAD :: T.S - 500 082

U.O.Note No.PL-II/A/27/2014/Rad.

Date: 18.08.2016.

Sub:- Procurement of equipment through e-procurement - Amendment to the technical specifications of certain equipment - Uploading in e-procurement.gov.in -  
Regarding.

Ref:-Tender Notification Rc.No.Plng-II/9/2016/TN, dt.23.6.2016.

-:o0o:-

The amendments to the technical specifications of **Computed Radiography Unit** and **300 mA X-Ray Machine** (Table 6 equipments) for displaying the same in e-procurement.gov.in.



Dr.N.Satyanarayana  
EXECUTIVE REGISTRAR

Amendment to the technical specifications of Computed Radiography Unit

S.No.	Tender Specification	Amendment
2.1	CR Reader should be able to process 65 image plates/hour or more of the largest size cassette.	CR Reader should be able to process 60 plates/hour or more @ 10 pixels/mm..
2.2	Resolution of 6 pixels/mm (minimum) for standard resolution cassettes and 10 pixels/mm (minimum) for high resolution cassette reading.	Resolution of 5 pixels/mm (minimum) for standard resolution cassettes and 10 pixels/mm (minimum) for high resolution cassette reading.
4.0	IMAGE ARCHIVAL - IMAGER/CAMERA	
4.1	Dry Chemistry laser thermal camera of minimum 500 DPI resolution.	Dry Chemistry laser/thermal camera of minimum 500 DPI resolution.

  
Executive Registrar

Amendment to the technical specifications of 300 mA X-Ray Machine

S.No.	Tender Specification	Amendment
2.3	Operating tube mA range should be 60 to 300 mA.	Operating tube mA range should be 60 to 300 mA or more.

  
Executive Registrar